

Sterilisation and Instructions for Use – Product Schedule

Diamond Brushes Right Angle

Classification: Class 2a Medical Device – (multi-use device)

GMDN Code	GMDN Term	Basic UDI-DI	Medical Device	MD Class
35768	Dental Professional Prophylaxis Brush, reusable	5051717DP01TU	JC, DIA Brushes	IIa
			KS, DIA Brushes	
			SET, 3022	

JC/DIA/A	Junior Cup, Dia Brush, Gold Cup, RA
KS/DIA/A/001	Dia Brush, Pencil, RA, Gold Cup
KS/DIA/A/002	Dia Brush, Pencil, Pointed, RA Gold Cup

RECOMMENDED SPEEDS

- Maximum speed 5,000 rpm
- Keep within the speed range of 1,500 – 5,000 rpm for the best work results and to increase the service life of the instrument.

INTENDED CLINICAL USE

The devices are indicated for use by qualified dental professionals to provide a highly versatile polishing system for all kinds of restorations materials.

Please refer to the Sterilisation and Instructions for Use for Multi-Use Devices for the correct procedures.