

Sterilisation and Instructions for Use – Product Schedule

SiCa Brushes Right Angle

Classification: Class 2a Medical Device – (multi-use device)

GMDN Code	GMDN Term	Basic UDI-DI	Medical Device	MD Class
35768	Dental Professional Prophylaxis Brush, reusable	5051717DP01TU	JC, SIC Brushes	IIa
			KS, SIC Brushes	
			SET, 3004	

JC/SICRA/B	Junior Cup, Silicon Carbide, RA
KS/SICRA/C	Silicon Carbide, Pencil, RA
KS/SICRA/D	Silicon Carbide, Pencil, Pointed, RA
JC/SICRA/G	Junior Cup, Silicon Carbide, Gold Cup, RA
KS/SICRA/E	Silicon Carbide, Pencil, Gold Cup, RA
KS/SICRA/F	Silicon Carbide, Pencil, Pointed, Gold Cup, RA

RECOMMENDED SPEEDS

- Maximum speed is 5,000 rpm.
- Keep within the speed range of 1,500 – 5,000 rpm for the best work results and to increase the service life of the instrument.

INTENDED CLINICAL USE

The devices are indicated for use by qualified dental professionals to provide a fast and effective polishing system for restorations and natural teeth.

Please refer to the Sterilisation and Instructions for Use for Multi-Use Devices for the correct procedures.